

This rejection is traversed and no longer applicable in view of the present amendment.

The Sancoff et al patents are disqualified as references because they and the subject invention were at the time the subject invention was made owned by the same person, Block Medical, Inc. The inventors were employed by Block Medical, Inc. at the time of making the invention, at the time of concurrently executing the declaration, and the assignment. The Sancoff et al patents were at the time of granting and are still assigned to Block Medical.


The Leibinsohn Patent U. S. 4,507,116 is directed to and discloses an apparatus for the induced infusion of liquid from a flexible bag wherein a flexible reservoir is pressurized by means of an inflatable bladder or "presser", both confined within a flexible sleeve. The bladder 6 of the bag presser 4 is inflatable by means of a pumping bulb 10 connected via tubing 12 and 13, and having an accumulator unit designated at 20.

The rejection cannot be sustained on the basis of Leibinsohn and Hessel '693. Accordingly, the rejection should be withdrawn and the application passed to issue and such action is respectfully solicited.

The objections to the drawings are noted, and new drawings will be provided as soon as the application is allowed. A photocopy of sheet 1 showing proposed corrections is enclosed.

No additional fees are due.




Respectfully submitted,



Freling E. Baker
Registration No. 24,078

BAKER, MAXHAM, JESTER & MEADOR
Symphony Towers
750 "B" Street, Suite 2770
San Diego, California 92101

Telephone: (619) 233-9004



Approved Drawing Convention
 8-23-93